

# MULTI GLOSS 268 Plus (20°, 60°, 85°) UNI GLOSS 60 Plus (60°)

### More comprehensive communication functions for more versatility.



## Large easy to read liquid crystal display. Easy operation with mode scroll wheel.

By operating the mode scroll wheel in accordance with easy-to-understand menu items, you can set a wide array of functions easily.



angle from the Geometry selection

screen and display two angles or


three angles at the same time.

(MULTI GLOSS 268Plus)

Wide measurement range that can measure from matte plastic to metal surface with a high gloss finish: 0.0~2,000 GU (in the case of 20°)

## Gloss difference measurement of standard/sample and PASS/FAIL evaluation.

You can perform a pass/fail evaluation by setting the limit value. You can enter the sample name and the standard value name in the measurement result.




#### **Autodiagnosis function for calibration**

Reduces needed frequency of calibration and improves long-term stability. In addition, if the calibration is required, the message will be displayed.



#### With dedicated Data transfer software (gloss-link), data management such as data transfer to Excel®\* is easy.




Microsoft Excel® 2000 or higher. Windows®and Excel® is a trademark or registered trademark of Microsoft Corporation of America or its subsidiaries.

#### **Specifications**

Model	<b>MULTI GLOSS 268Plus</b>	UNI GLOSS 60Plus
Measurement geometry	20°, 60°, 85°	60°
Size of measurement	20°: 10 X 10 mm	
spot	60°: 9 X 15 mm	60°: 9 X 15 mm
	85°: 5 X 38 mm	
Standard compliance	ISO 2813, ISO 7668, ASTM D 523,	
	ASTM D 2457, DIN 67 530, JIS Z 8741,	
	BS 3900, BS 6161(Part 12)	
Measurement	20°: 0.0~2,000 GU	
range	60°: 0.0~1,000 GU	60°: 0.0~1,000 GU
	85°: 0.0~160 GU	
Resolution	0.0~99.9 GU : 0.1 GU	
	100~2,000 GU : 1 GU	
Repeatability	0.0~99.9 GU : 0.2 GU	
	100~2,000 GU : 0.2% of reading	
Inter instrument	0.0~99.9 GU : 0.5 GU	
agreement	100~2,000 GU : 0.5% of reading	
Memory	999 measurements with date and time	
Difference measurement	Memory for 50 standard	
Battery performance	Approx. 10,000 measurements	
	(when using 1.5 V AA(R6)size alkaline battery)	
Measurement time	0.5 seconds	
Auto power off	10~99 seconds selectable	
Languages	English, Spanish, German, French, Italian, Japanese,	
	Portuguese, Russian, Polish	
Measurement mode	Normal mode (Sample mode, Statistics, Continuous,	
	Basic mode), Difference mode	
Interface	USB2.0; Bluetooth® standard version 2.0*	
Operating temperature	15 to 40°C	
/humidity range	relative humidity 85% or less (at 35°C) with no condensation	
Storage temperature	-10 to 60°C	
/humidity range	relative humidity 85% or less (at 35°C) with no condensation	
Power supply	1.5 V AA(R6)size alkaline battery, via USB port	
Dimensions	155(W) X 73(H) X 48(D) mm	
Weight	400 g	

Note: The specifications described in this document are subject to change without notice.

#### **Dimensions**



Units : mm

#### **Accessories**

Model	MULTI GLOSS 268Plus	UNI GLOSS 60Plus	
Standard	Calibration holder TRI	Calibration holder 60°	
accessories	(High gloss tile)	(High gloss tile)	
	USB-cable		
	Carrying case		
	Data transfer software (gloss-link)		
	1.5 V AA(R6)size alkaline battery		
Optional	Checking standard TRI	Checking standard 60°	
accessories	(3 semi gloss tiles 20, 60, 85°)	(High and semi gloss tile)	
	Checking standard mirror		
	(3 tiles 20, 60, 85°)		

\* Applicable Bluetooth® profile: Serial Port Profile, Output: Bluetooth® Power Class 1 The communication distance may vary depending on the obstacles and radio wave conditions between the devices. Successful wireless communication is not guaranteed with all Bluetooth®-ready equipment.

MULTI GLOSS 268 / UNI GLOSS 60 conforms with international specifications ASTM D 523, DIN 67 530, ISO 2813.











Certificate No : LRQ 0960094/A
Registration Date : March 3, 1995
Registration Date : March 12, 1997

#### SAFETY PRECAUTIONS

For correct use and for your safety, be sure to read the instruction manual before using the instrument.

- Always connect the instrument to the specified power supply voltage. Improper connection may cause a fire or electric shock.
- Be sure to use the specified batteries. Using improper batteries may cause a fire or electric shock.

KONICA MINOLTA OPTICS, INC. Konica Minolta Sensing Americas, Inc Konica Minolta Sensing Europe B.V.

Konica Minolta (CHINA) Investment Ltd. SE Sales Division
Beijing Office
Guangzhou Office
Chongqing Office
Qingdao Office
Wuhan Office

Konica Minolta Optics, Inc.

Thailand Representative Office Bangkok, Thailand

Addresses and telephone/fax numbers are subject to change without notice. For the latest contact information, please refer to the KONICA MINOLTA OPTICS Worldwide Offices web page:

Thailand Representative Office Bangkok, Thailand

Thoiland

Phone: +66-2361-3730

Fax: +66-2361-3771

http://konicaminolta.com/instruments/about/network

Konica Minolta Sensing Singapore Pte Ltd. Konica Minolta Optics, Inc. Korea

<sup>\*</sup>Bluetooth® is a registered trademark of Bluetooth SIG, Inc.